

26/ Feb / 05

| | | | |
|-------|----------|--------------------------------|---|
| 13:19 | Run 8952 | Pedestal, circulation pump on | |
| 13:21 | Run 8953 | LED, circulation pump on | |
| 13:27 | Run 8954 | α , circulation pump on | (ish 97m) $\lambda_{abs} = 110.8 \text{ cm}$ |
| 14:21 | Run 8955 | α , " | (ish 94m) $\lambda_{abs} = 125.5 \text{ cm}$ |
| 15:26 | Run 8956 | α , " | $\lambda_{abs} = 118.1 \text{ cm}$ |
| 16:21 | Run 8957 | α , " | $\lambda_{abs} = 119.9 \text{ cm}$ |
| 17:26 | Run 8958 | α , " | $\lambda_{abs} = 120.5$ |

Increased threshold of discriminating α from ~~75~~ to ~~125~~ mV

| | | | |
|-------|-------|----------|----------|
| 18:17 | #8959 | α | 106.8 cm |
| 19:33 | #8960 | α | 111 cm |
| 20:53 | #8961 | α | 107.1 cm |
| 21:12 | #8962 | pedestal | |
| 21:14 | #8963 | LED | (Junk) |


New steering parameters of ROME Task XeAlphaFis was introduced

These parameters can be changed with editing "rome Config.xml"

In Liquid : switch Liquid/Gas

CircleFit : switch whether circle fit ~~will be~~ ^{is} done.

AbsorptionLengthFit / Min, Max : range of expo fit



Y, Z : center of α rings, these values will be used if CircleFit is false

P.S. there is a switch to enable/disable Canvas drawing.

Calibration file updated

| | | | |
|-------|-------|----------|---------|
| 22:13 | #8964 | α | 96.9 cm |
| 23:26 | #8965 | LED | |
| 23:40 | | | |

Stop liquid phase circulation

- close valves between chamber and purification line. (getter)
- keep opening valves between chamber and liquid pump cartridge

Calibration file updated (only pedestal)

| | | | |
|-------------|-------|----------------|--|
| 23:44 | #8966 | α | 98.44 cm |
| | #8967 | pedestal | |
| 27/Feb/2005 | | | |
| 1:05 | #8968 | α | (after ~1 hour from pump stopped). $\lambda_{abs} = 92.06 \text{ cm}$ |
| 3:00 | #8969 | α run, | after 3 hours from pump stopped. $\lambda_{abs} = 95.3 \text{ cm}$ |
| 4:00 | #8970 | α run | after 4 hours from pump stopped. $\lambda_{abs} = 99.16 \text{ cm}$ |
| 5:00 | #8971 | α run, | after 5 hours from pump stopped. $\lambda_{abs} = 99.12 \text{ cm}$ |
| 6:00 | #8972 | α run = | after 6 hours from pump stopped, $\lambda_{abs} = 95.21 \text{ cm}$ |


27/02/2005.

6:13 RUN #8923. pedestal run.

RUN #8924. LED calibration run.

6:27 RUN #8925. Cosmic Ray. triggered by TCs.
@ pump off.

■ TOTAL PLOT @ YESTERDAY and THIS MORNING.



RUN #8932
?
RUN #8922.

● : yesterday morning.
(PUMP ON).
RUN 8932 ~ 8945 (~8 hour).

■ : yesterday daytime.
(PUMP ON).
RUN 8950 ~ 8964.

▲ : this morning.
(PUMP OFF).
RUN 8966 ~ 8972.

- According to, yesterday morning's data (●), λ_{abs} continued to grow.
- According to, yesterday daytime's data (■), λ_{abs} continued to grow by ~4 hours, but it looks like saturated (???) and after 7 hours, it continued to grow down.
- According to pump off data (▲), it looks like there is no time evolution.

- Now, the valve b/w LP and liquid pump cartridge are closed, so this cartridge has not affected the λ_{abs} . not.

- $\langle \lambda_{abs} \rangle = 96.63 \text{ cm. } (\pm 2.56 \text{ cm, RMS})$
(pump off) #8966 ~ #8972.

- This is current offset of absorption length.

7:48 #8925, paused. due to check the trigger logic and TDC timing because TDC distribution is strange @ LP monitor. TDC entry of telescope is stored by only TC2-bottom. why? I checked trigger logic. and delay timing. But it seems correct, no problem....

8:23 #8925 resumed.

8:34 #8925. (cosmic). Stopped.


8:52 #8926. pedestal. @ pump off.

8:53 #8922. LED @ pump off. Calibration file updated

11:18 #8978 α , pump off. $\lambda_{abs} = 89.0 \text{ cm}$

12:35 MSCB Access Error
- LabVIEW restarted
- MSCB crate rebooted

15:49 #8979 α , pump off. $\lambda_{abs} = 116.5 \text{ cm}$



27 Feb 05

15:56 #8980 CR

18:29 Stop the Run 291 events

18:30 #8981 pedestal

18:37 #8982 LED

18:37 #8983 α $\lambda_{abs} = 110.0 \text{ cm}$ Calibration file updated

19:27 #8984 α $\lambda_{abs} = 115.8 \text{ cm}$

20:27 #8985 α $\lambda_{abs} = 108.2 \text{ cm}$

21:27 #8986 α $\lambda_{abs} = 110.5 \text{ cm}$

21:45 lig. pump started cartridge temp -75°C

21:51 reached 53 Hz (53.09 Hz) " -92°C

21:53 Purifier Pressure Range Setting 0.147 - 0.150 MPa
21:56 Purifier pressure reached $\sim 0.27 \text{ MPa}$

22:00 -96°C

22:02 -97°C

22:02 #8987 Junk RUN $\lambda_{abs} = 116.1 \text{ cm}$

22:02 #8988 α $\lambda_{abs} = 116.1 \text{ cm}$

22:11 Pressure Range Setting 0.147 - 0.152 MPa

22:12 cartridge temp -99°C


22:08 #8989 pedestal

22:10 #8990 LED

22:16 #8991 α again $\lambda_{abs} = 111.9 \text{ cm}$

Please continue this circulation for at least 24 hours!

27 Feb 05




27 Feb 05

0:00

23:03 #8992 α $\lambda_{abs} = 115.8 \text{ cm}$

23:13 #8993 CR Stopped at 0:04 91 events
23:24 Pressure Range Setting 0.148 - 0.152



28/ Feb 05

28/02/05

0:04 #8994 α $\lambda_{abs} = 103.8$

0:11 #8995 CR

2:02 #8995 stop

#8996 α $\lambda_{abs} = 114.4 \text{ cm}$

2:10 #8997 CR

4:00 #8997 stop

#8998 α $\lambda_{abs} = 107.0 \text{ cm}$

#8999 CR

6:00 #8999 stop

#9000 α $\lambda_{abs} = 117.4 \text{ cm}$

#9001 CR

7:40 #9001 stop

#9002 pedestal

#9003 LED

8:00 #9004 Alpha @ pump ON. (~ 10 hours) $\lambda_{abs} = 109.9 \text{ cm}$

8:08 #9005 CR

9:54 #9005 stop

9:52 #9006 Alpha @ ~ 12 hours. $\lambda_{abs} = 119.4 \text{ cm}$

10:00 #9007 CR @ pump on.

11:45 #9008 Alpha @ ~ 14 hours. $\lambda_{abs} = 119.7 \text{ cm}$

#9009 CR @ pump on (~ 14 hours)

14:00 #9010 α ~ 16 h $\lambda_{abs} = 126.2 \text{ cm}$

14:07 #9011 CR pump on

15:46 #9012 pedestal pump on

15:48 #9013 LED pump on

15:53 #9014 α ~ 18 h $\lambda_{abs} = 128.4 \text{ cm}$

16:25 #9015 LED pump on

16:41 #9016 CR

17:49 end 104 events

17:49 #9017 alpha ~ 20

17:57 #9018 CR $\lambda_{abs} = 97.22 \text{ cm}$

19:43 end 175 events

19:44 #9019 α ~~175~~ ~ 22 hours after pump started

19:50 #9020 CR 103 events $\lambda_{abs} = 124 \text{ cm}$

20:45 #9021 α $\lambda_{abs} = 120.5 \text{ cm}$

20:52 #9022 CR

21:44 #9023 α $\lambda_{abs} = 109.2 \text{ cm}$

21:58 #9024 pedestal

21:59 #9025 LED

23:30 #9026 α

0:59 #9027 α

2:58 #9028 ~~CR~~ α

#9029 CR

4:19 #9029 stop

#9030 pedestal

#9031 LED

01/03/05

Using this, $\lambda < 0$

Using #9012, #9013, $\lambda \sim 100$ $\lambda_{abs} = 121.9 \text{ cm}$

$\lambda_{abs} = 157.7$ (ped #9012, LED) #9026

$\lambda = 115.7$ (ped, LED: #9012, #9013) 123.6 (ped, LED: #9030, #9031)

Calibration change

1/Mar/05

5:00 #9032 α $\lambda_{abs} = 122.5 \text{ cm}$
 #9033 CR
 7:00 end 196 events
 #9034 α $\lambda_{abs} = 116.6 \text{ cm}$
 #9035 CR 227 ev.
 9:07 #9036 pedestal
 9:10 #9037 LED
 9:17 #9038 alpha @ pump ON (~ 35 hours)
 9:27 #9039 CR $\lambda_{abs} = 129.8 \text{ cm}$
 ~140 ev.

SC frontend restarted because of the message.
 "Error reading MSCB bus"

10:53 #9040 alpha $\lambda_{abs} = 119.5 \text{ cm}$
 11:10 #9041 CR
 14:54 #9042 pedestal
 14:55 #9043 LED
 15:01 #9044 alpha $\lambda_{abs} = 104.9 \text{ cm}$
 15:11 #9045 CR
 17:01 #9046 alpha $\lambda_{abs} = 112.4 \text{ cm}$
 17:07 #9047 CR
 17:26 CE tank 4100 l LN₂
 18:58 #9048 alpha $\lambda_{abs} = 72.21 \text{ cm}$ (91.48 cm)
 19:05 #9049 CR
 ↑
 Something wrong!?
 Take once more

1/Mar/05

20:06 #9050 alpha $\lambda_{abs} \approx 90 \text{ cm}$
 too short
 Take calibration data again
 20:53 #9051 pedestal
 20:54 #9052 LED
 21:03 #9053 α $\lambda_{abs} = 25.9 \text{ cm} !?$

HV trouble when I realized the trouble

~~21:20~~ 22 HV off.

Network access was impossible
 Monitor was black even if I pushed button.
~~Light~~ Light corresponding slot { 0, 2 - off
 1, 3 - on
 I couldn't turn off HV when I changed mode to local and ~~the~~ pressed "HV off" button.

So I pressed "Panic off" button and turned key to off

Midas message
 17:29 Connection broken to LRS1454 from SC frontend

After that HV02 became unable to start. Then I cleaned inside with air flow but it didn't work. I ~~decided~~ decided to continue data acquisition tonight, and began to fix problem tomorrow.

23:32 #9055 pedestal without hv02
 23:33 #9056 LED "
 23:40 #9057 α "

2/Mar/05

01:45 #9058 α "
 #9059 ← ← Junk hv01 was powered off

2/Mar/05

| | | | |
|------|-------|----------|------------|
| 8:38 | #9060 | pedestal | HV 02 off. |
| 8:40 | #9061 | LED | " |
| 8:46 | #9062 | alpha | " |
| 9:02 | #9063 | CR | " |

9:10 MSCB Access error. → reboot.

19:40 HV 02 recovered HV on

| | | |
|-------|-------|----------|
| 20:40 | #9067 | pedestal |
| 20:41 | #9068 | LED |
| 20:47 | #9069 | α |
| 21:27 | #9070 | LED Junk |
| 21:28 | #9071 | LED |

No signal (@ ADC data) PMT

F23
F24
F35
F34
F6
F0
T40
BK33
BK34
BT37
BK31
BK23

No signal at input of splitter


signal at output of splitter O.K.

BNC connector to splitter input was broken To be repaired

replaced APC mini card for BK33, BK34, BT37, BK31, T40

↓
APC count O.K.

F23 supply 1100 V from HV02-0-0, LED on
77 μA strange signal (at input of splitter)



it seems like cross talk but ~~is~~ too high.

Breeder circuit might be broken

F24 ~~No signal~~ same as F23
 F35 " "
 F34 same as F23 but smaller signal
 F6 same as F23
 F0 "

BNC Connector of BK23 was repaired

Summary

- Strange signal PMTs F23, F24, F35, F34, F6, F0
Probably, circuit was broken when HV trouble occurred.

← See next to next page.

- APC mini cards T40 BK33 BK34 BT37 BK31 ⇒ Fixed
- BNC connector BK23 ⇒ Fixed

| | | | |
|-------------|-------|----------|---|
| 22:41 | #9072 | LED | |
| 22:44 | #9073 | " | |
| 03/Mar/2005 | 0:02 | #9074 | α |
| 0:03 | #9075 | α | |
| 0:26 | #9076 | CR | |
| 1:19 | #9077 | pedestal | |
| 1:20 | #9078 | LED | |

→ see next page

24 / Mar / 2005

1:36 #9079 CR
 2:00 #9080 ~~CR~~ α
 #9081 CR
 4:00 #9082 α
 #9083 CR
 6:00 #9084 α

8:00 all HT had been off. \rightarrow reload. 050220-2.lv file.
 since #9072 ??

8:20 #9085 pedestal
 8:22 #9086 LED
 8:31 #9087 alpha
 8:39 #9088 CR \rightarrow 0 events

Burndy SG-16 disconnected !! \rightarrow fixed.

9:01 #9089 pedestal
 9:03 #9090 LED
 9:09 #9091 alpha $\lambda_{abs} = 123.7$ cm.
 9:17 #9092 CR \rightarrow 1 events.
 trigger cable for CR disconnected \rightarrow fixed.
 9:35 #9093 alpha $\lambda_{abs} = 119.3$ cm
 9:44 #9094 CR 37 events.
 10:08 #9095 alpha $\lambda_{abs} = 114.5$ cm

23 / Mar / 2005

10:22 #9096 pedestal
 10:24 #9097 LED
 10:33 #9098 alpha $\lambda_{abs} = 133.8$ cm

10:41 #9099 CR
 11:51 #9100 alpha $\lambda_{abs} = 130.7$ cm.

11:59 #9101 pedestal.
 12:00 #9102 LED

12:09 #9103 alpha $\lambda_{abs} = 128.6$ cm
 12:18 #9104 CR
 14:01 #9105 alpha $\lambda_{abs} = 139.2$ cm

14:11 #9106 pedestal
 14:13 #9107 LED
 14:20 #9108 alpha $\lambda_{abs} = 130.1$ cm.
 14:29 #9109 CR

Signal cables at feedthrough almost disconnected. (#4 board)
 L19, L22, L23, T19, T23, R22, R23, BT19, BT22, BT23. no signal.
 \rightarrow fixed.
 at the same time, F23, F24, F35, F34, F6, F0 seems O.K.

~~14:29~~
 15:24 #9110 pedestal
 15:27 #9111 LED
 15:34 #9112 alpha $\lambda_{abs} = 117.1$ cm
 15:42 #9113 CR
 17:38 #9114 pedestal
 17:39 #9115 α $\lambda_{abs} = 107.5$ cm
 18:54 #9116 α $\lambda_{abs} = 112.8$ cm


pressure control setting $\frac{1.46}{1.50} \rightarrow \frac{1.49}{1.53} \times 10^{-1}$ Mpa

03/Mar/2005

19:52 #9117 α $\lambda_{abs} = 106.9$ cm
 21:35 #9118 α $\lambda_{abs} = 111.9$ cm
 23:01 #9119 α $\lambda_{abs} = 111.9$ cm

04/Mar/2005


0:01 #9120 pedestal
 0:01 #9121 LED
 0:30 #9122 α $\lambda_{abs} = 116.3$ cm



Analysis of #9108 with various correction for each α source


this analysis code can be used under lp framework.new which shows existing plot for α currently, reference is MC

data/~~#9122~~ #8528 w/o correction




48 117
 87 137 cm

data/#8528



~~47.6 87.7~~
~~117 136.6 cm~~
 49 109
 98 120 cm
 corrected with distance from source $\cos(\theta)$


data/#8528 corrected with correction factor



47 145
 115 124 cm

Correction factor was estimated from $MC(\text{current } \alpha \text{ position}) \div MC(\text{former } \alpha \text{ position})$

data/MC



112 647
 435 1266 cm

From this result, assumed position may not be correct.

4/ Mar, 2005

| | | | |
|------|-------|---|------------------------------------|
| 2:00 | #9123 | α | $\lambda_{abs} = 110.6 \text{ cm}$ |
| 4:12 | #9124 | α | $\lambda_{abs} = 104.0 \text{ cm}$ |
| 6:00 | #9125 | α | $\lambda_{abs} = 93.6$ |
| 8:10 | #9126 | α | $\lambda_{abs} = 121 \text{ cm}$ |

← During this Run Message Connection broken to LRS#54 LRS1454 doesn't respond

7:56 #9127 pedestal

7:58 #9128 LED

8:03 #9129 α $\lambda_{abs} = 119.6 \text{ cm}$

dpf/beam work hung up. All operation on meganucol ~~was~~ impossible. meganucol was rebooted manually.

8:46 meganucol recovered

8:48 #9130 CR
seems to work.

10:01 #9131 α $\lambda_{abs} = 128.2 \text{ cm}$

10:08 #9132 CR

12:07 #9133 α $\lambda_{abs} = 125.1 \text{ cm}$

12:16 #9134 CR

14:00 Slow Control was found to be stopped. Restarted

14:03 #9135 α

Stopped intermediate because of SC hung

14:09 #9136 α restarted & failed due to meganucol hung up.

04 (Mar/2005)

meganucol rebooted manually.

14:52 #9136 again α.

After several iterations, pressure range setting is determined to be 0.149 ~ 0.153 MPa for stable liquid circulation

17:03 ~~pressure range 1.51 - 1.56~~

17:21 • Pressure control range 1.53 - 1.56 $\times 10^{-1}$ MPa
• changed interval of reading MSCB and control LN₂ value from 1 sec to 200 msec.

Now circulation pump is stable.

17:24 Get Pressure control range back ~~(0.149 - 0.153 MPa)~~
(0.149 - 0.153 MPa)

It seems OK.

17:25 #9137 pedestal

#9138 LED

17:40 #9139 α 133.6 cm

decreased MIM Discriminator level for sum out of Discriminator 0 (front top left α) from 125 to 75 mV

18:27 #9140 α 129.8 cm

18:57 #9141 α 153.3 cm

This is for removing possible trigger bias which is thought to cause "excess" in short distance range of Front-Left-Top histogram.

19:05
19:12

stop liquid phase circulation
start gas phase circulation

pressure range 0.132 - 0.136 MPa
Flow rate - 7.2 ~ 8.3 l/min gas

19:57 #9142 α 111.5 cm
21:07 #9143 α Junk. 141.3 cm

megacoin01 kernel panic during #9143

21:21 #9144 α 113.7 cm
~~21:50~~ #9145 α 110.1 cm
#9146 Junk
23:38 #9147 α 114.2 cm

05/Mar/05 (Sat.)

1:30 #9148 α 118 cm
3:27 #9149 pedestal Junk
#9150 LED Junk
3:35 #9151 pedestal Junk

SCFrontend started but soon stopped I tried again & again

I found LabVIEW frozen, forced termination and restart. but. it ~~does~~ not respond. so, reboot MSCB. → OK.

start SCFrontend → OK.

since 3:06, LabVIEW have been frozen. 3:07 SCFrontend was stopped.

~~3:55~~ 4:03 restart LabVIEW Run. inner pressure up to 0.26 MPa. now flow rate is 0.

4:17 #9152 Pedestal
4:18 #9153 LED
4:31 #9154 α λ_{obs} = 105.7 cm
5:33 #9155 α λ_{obs} = 114.8 cm
7:00 #9156 α λ_{obs} = 121.8 cm

7:30. Gas phase circulation pump restart. Flow rate 7.3 ~ 8. l/min

pump has been off since the slow control trouble

7:47 #9157 CR
9:16 #9158 α λ_{obs} = 106.6 cm
9:24 #9159 pedestal
9:26 #9150 LED
9:32 #9161 CR

No data in CR. trigger ??? From when? Stop this RUN and try to take CR. self trigger.

05/Mar/05

11:44 CR trigger was changed from Counter-trigger to self-trigger

11:41 # 9162 CR (self) OK. data is in. Probably timing is incorrect in CR-counter trigger.

12:56 # 9163 α $\lambda_{abs} = 106.2 \text{ cm}$

Check the COSMIC RAY TRIGGER counter timing.
→ Timing seems to be OK. (ADC timing)

13:50 # 9164 CR
Trigger by upper trigger counters.

→ Found that TBC inputs of TC2 up & down are swapped & fixed.

14:07 # 9165 CR
Trigger by lower trigger counters

14:12 # 9166 CR
Trigger by ~~TC~~ TC2 lower.

14:22 # 9167 CR
Trigger by TC2 upper

14:33 # 9168 CR
Trigger by TC1 upper

14:56 # 9169 CR
Trigger by TC1 lower

15:06 # 9170 CR
Trigger by TC3 upper

15:13 # 9171 CR
Trigger by TC3 lower

05/Mar/05

15:20 # 9172 α Data quality seems to be bad? because we cannot see clear rings.

Take calibration data again and then d.

15:23 # 9173 pedestal

15:25 # 9174 LED

15:38 # 9175 α again $\lambda_{abs} = 102.3 \text{ cm}$

Now CR trigger is fired by TRIGGER Counters.

16:07 # 9176 CR

FROM online histograms in RUN #9166~9177, the counters are working more or less. although some of them are very noisy. The threshold levels seem to be rather low - resulting in quite many random coincidence fake trigger... Please be careful when you look at the data.

16:30 AmBe source in use in front of the detector

17:32 # 9177 α $\lambda_{abs} = 99.36 \text{ cm}$

17:43 # 9178 CR

19:25 # 9179 α $\lambda_{abs} = 85.99 \text{ cm}$

05/Mar/05


19:32 #9180 CR
 21:40 #9181 α $\lambda_{abs} = 89.07 \text{ cm}$
 21:47 #9182 CR
 23:08 #9183 α $\lambda_{abs} = 104.3 \text{ cm}$
 23:15 #9184 CR

06/Mar./2005, (Sun.)

0:48 #9185 pedestal
 0:50 #9186 LED
 1:00 #9187 α $\lambda_{abs} = 110.1 \text{ cm}$
 2:04 #9188 CR
 3:00 #9189 α $\lambda_{abs} = 112.6 \text{ cm}$
 3:08 #9190 CR
 5:02 #9191 α $\lambda_{abs} = 114.4 \text{ cm}$
 5:33 #9192 CR
 7:00 #9193 α $\lambda_{abs} = 115.7 \text{ cm}$
 7:11 #9194 CR

9:59 end of #9194

Implement gamma trigger for Ni- γ by using central 12 PMTs on the front face.




06/Mar/05

14:03 #9195 α mistake in trigger.
Band data

11:08 #9196 α $\lambda_{abs} = 129.1 \text{ cm}$

- Am-Be source in front of the LP.
- Window cover of the LP removed.



11:23 #9197 γ trigger for Ni- γ w/ Am-Be/Ni

11:59 #9198 γ trigger w/o Am-Be/Ni

Analogue SUM. changed. Now Only using F14, F15, F20 and F21

12:22 #9199 γ trigger w/ Am-Be/Ni


12:40 #9200 Pedestal.

12:41 #9201 LED

12:53 #9202 CR

14:28 # 9203 α $\lambda_{abs} = 735.6 \text{ cm}$


Analysis # 9199



gsum vs
 $ADC(F14)$
 $ADC(F15)$
 $ADC(F20)$
 $ADC(F21)$

gsum.
 $ADC(F14) + \dots + ADC(F21) > 750$

15:06 # 9204
 γ trigger w/ Am-Be/Ni
 Triggered by central 4PMTs
 F14, F15, F20, F21



Almost same charge
 but
 different pulse height

16:30 Ni plates 2.5mm x 4 removed

16:34 # 9205
 γ trigger w/ Am-Be && w/o Ni plates

17:45 Run 9205 stopped after 237657 events.
 Due to a problem with the data logger this file may be corrupted. Hence we started a new identical run

9206 w/ Am Be w/o Ni Plates

↑
 20 km/sec
 SOURCE

18:45 " stopped

9207 Pedestal 5k events

9208 LED calibration

9209 Am Be + 1 cm Nickel

19:45 " stopped after 900k events

20:10 # 9210 α

20:18 # 9211 CR

22:03 # 9212 α

$\lambda_{abs} = 103.7 \text{ cm}$

22:09 # 9213 CR

07 / March / 2005

0:05 # 9214 pedestal

0:08 # 9215 LED

0:17 # 9216 alpha

$\lambda_{abs} = 110.4 \text{ cm}$

0:27 # 9217 CR

2:07 # 9218 alpha

$\lambda_{abs} = 106 \text{ cm}$

2:13 # 9219 CR

4:20 meganlnc1 hang up \rightarrow reboot.

07/ March/2005

4:30 #9220 pedestal
 4:32 #9221 LED
 4:44 #9222 alpha $\lambda_{abs} = 87 \text{ cm}$
 5:10 #9223 CR
 6:06 #9224 alpha $\lambda_{abs} = 97.71 \text{ cm}$
 SC frontend was abnormally stopped \rightarrow restarted.

6:14 #9225 alpha $\lambda_{abs} = 98.7 \text{ cm}$

MSCB frozen \rightarrow restarted

circulation pump is not working. too hot? flow rate ~ 0

\rightarrow fixed by Rye.

AC 100V voltage shortage \rightarrow line rearranged.

7:45 #9226 CR
 8:07 #9227 pedestal
 8:09 #9228 LED
 8:15 #9229 α $\lambda_{abs} = 91.7 \text{ cm}$
 8:31 #9230 CR
 10:45 stopped


#9231 Pedestal

#9232 LED

- We increased the discrimination threshold on the linear sum of the front 4 central PMTs in order to get rid of the α -background


- we use the 2km/sec source (health care people took the high intensity one)

#9233 low intensity AuBe source, nichel plate present



12:00 Run stopped $\sim 60 \text{ k}$ events

#9234 same as run 9233 BUT W/O NICKEL



13:35 RUN 9234 stopped $\sim 88 \text{ k}$ events

RUN 9235 SAME AS 9234 But with high intensity SOURCE

14:35 " stopped.

Removed the nickel in front of NaI.

14:40 RUN 9236 same as 9235 (Au/Be HIGH int. NO NICKEL)

15:40 stopped run 9236 @ 100k events.

RUN 9236 Au/Be high intensity with 1cm nickel plate (see fig 1 without nickel on NaI)

16:50 Run 9237 stopped ~100k events
 17:30 # 9238 same as 9237
 18:35 " stopped
 # 9239 same as run 9236 (Ambx high No nickel)
 19:20 Lab view frozen
 SC Frontend stoped.
 ↑
 ~19:40 I found it at ~19:40 so
 Reboot MscB → OK
 Restart LabView → OK
 Start SC Front End → OK

9239 Stop


19:52 # 9240 α λ_{abs} = 107.6 cm
 19:59 # 9241 CR
 22:01 # 9242 α λ_{abs} = 105.6 cm
 22:09 # 9243 CR

8/March/2005

0:13 #9244 pedestal
 0:14 #9245 LED
 0:25 #9246 alpha λ_{abs} = 104.2 cm
 0:35 #9247 CR
 2:19 #9248 alpha
 2:29 #9249 CR λ_{abs} = 108.6 cm

08/March/2005

4:15 # 9250 pedestal
 4:17 # 9251 LED
 4:23 # 9252 alpha λ_{abs} = 105 cm
 4:31 # 9253 CR
 6:18 # 9254 alpha λ_{abs} = 95.7 cm
 6:30 # 9255 CR
 9:30 # " stopped
 9:30 # 9256 pedestal run
 # 9257 LED calibration run
 # 9258 LP Am/Be w/ nickel in this config:





This is the configuration that with the NeI gives the best resolution.

10:35 stopped
 10:48 #9259 LP Am/Be w/o nickel (normalization for 9258)
 12:00 " stopped.

RUNS 9236/7/8/9

Ni Line in Large Prototype



8/ Mar / 2005

12:18 #9260

pedestal

12:21 #9261

LED



12:36 #9262

α

$\lambda_{\text{obs}} = 111.9 \text{ cm}$

RUNS 9258/9259

Ni Line in Large Prototype



8/Mar/2005

14:15 #9264 pedestal
14:16 #9265 α $\lambda_{abs} = 77.63$

14:40

start liquid phase circulation

adjustment pressure setting
0.139 - 0.142 MPa

16:14 #9266 Pedestal circulation liquid pump on
16:15 #9267 LED
16:21 #9268 α $\lambda_{abs} = 97.5$ cm
16:29 #9269 CR
18:39 #9270 α $\lambda_{abs} = 113.7$ cm
18:47 #9271 CR
20:29 #9272 α $\lambda_{abs} = 123.8$ cm
20:37 #9273 CR
22:32 #9274 α $\lambda_{abs} = 128.3$ cm
22:40 #9275 CR

9/Mar./2005 (Wed)

00:22 #9276 α
00:29 #9277 CR $\lambda_{abs} = 135.6$ cm

2:13 #9278 pedestal
2:16 #9279 LED
2:27 #9280 alpha $\lambda_{abs} = 116$ cm

2:52 MSCB error \rightarrow reboot MSCB

#9281

3:20 pedestal test

3:59 alpha $\lambda_{abs} = 133.8$ cm

#9282

4:26 #9283 CR

6:29 #9284 alpha $\lambda_{abs} = 131.5$ cm.

CR ADC, TDC gate & timing re-checked.
Everything seems o.k. After that, on-line histogram is also o.k.

7:25 #9285 CR.

8:19 stop 9285

8:19 #9286 α $\lambda_{abs} = 136.5$ cm

8:30

stop liquid phase circulation


pressure control range. $1.28 - 1.33 \times 10^{-1}$ MPa

10:01 #9287 pedestal

10:06 #9288 LED

10:12 #9289 α $\lambda_{abs} = 117.8$ cm

Installation of Nitrogen target.



9 Mar 05

10:46

#9290

AmBe

(Front 4PMT Sum th = 100 mV
Bonner sphere dia = 5 inch

11:45

#9291

AmBe

(4PMT Sum th = 150 mV
Bonner sphere dia = 8 inch

changed detector from normal (maybe sus) to plastic (10.5 cm ϕ) beaker

Bonner sphere 8 inch \rightarrow 5 inch

12:12

#9292

AmBe

#9293

crash logger

14:45

#9294

AmBe

5 inch sphere

logger crashed before the end of RUN

RUN #9294 stopped

Nitrogen target removed


15:21

#9295

AmBe

5 inch sphere

w/o Nitrogen



16:10


#9296

AmBe

2 inch sphere

changed direction of sphere.

some $\square \odot$




16:31

#9297

AmBe

2 inch sphere

polyethylene w/ Boron



16:52

#9298

AmBe

2 inch sphere

Al target instead of LN2 (20 mm t)

17:13

#9299

AmBe

2 inch sphere

Al target 8 mm thickness

The 2-inch sphere is rotated by 90°


17:55

#9300

AmBe

2 inch sphere

effectively ~ 4cm thick moderator




Bonner Sphere

2 inch \rightarrow 8 inch

9 / Mar / 2005

18:13 #9304 AuBe
8 inch sphere.
2x4mm Al

Date disappeared !!!
Erased by the logger !!!



18:48 #9302 AuBe
8 inch sphere
20mm Al

Target changed from Al to LN₂

19:20 #9303 AuBe
8 inch sphere.
LN₂ target.

21:20 AuBe source removed.

(19:38 #9304
19:56 #9305

| | | | | |
|-------|-------|----------|-----------------------------|---------------|
| 21:31 | #9306 | Pedestal | Both calculation off. | |
| 21:32 | #9307 | LED | | |
| 21:39 | #9308 | α | λ _{abs} = 135.6 cm | (#9309 debug) |
| 23:34 | #9309 | α | λ _{abs} = 177.8 cm | |

COSMIC RAY TRIGGER MODIFIED

Previous $(T1 \cap B1) \cup (T2 \cap B2) \cup (T3 \cap B3)$

New $(T1 \cup T2 \cup T3) \cap (B1 \cup B2 \cup B3)$

1008.

| | | | |
|-------|-------|----------|-----------------------------|
| 1:05 | #9311 | CR | (new logic) |
| 6:43 | #9317 | end. | |
| 6:44 | #9312 | pedestal | |
| 6:44 | #9313 | LED | |
| 6:50 | #9314 | d. | λ _{abs} = 14.3 cm |
| 6:58 | #9315 | CR | (new logic) same as #9311 |
| 10:59 | #9316 | pedestal | |
| 11:18 | #9317 | pedestal | |
| 11:20 | #9318 | LED | |
| 11:27 | #9319 | α | λ _{abs} = 130.6 cm |
| 11:39 | #9320 | CR | , same as #9311 |
| 16:12 | #9321 | α | λ _{abs} = 116 cm |
| 16:24 | #9322 | CR | , same as #9311. |

17:20

~~17:20~~ start preparation to recovery

17:31

9323 $\alpha \leftarrow$ Junk

18:00

megonh01 hung up \rightarrow reboot.

18:14

9324 α

18:23

9325 CR, same as #9311

19:05


stop 9325

HV off

Recovering start

- Fill outer vessel with 0.25 atm N_2
- excite minco heater at 30V
(this is the maximum voltage)
- Don't forget evacuate outer vessel before inner pressure becomes lower than ~~1~~ 0.25 atm

SCS-700 (~~MSCB~~ module for PT100) is broken. I prepared a new VI which does not access this module. (meg-lpt-notmp.vi)
One of valves in recovering line is not fully open



Level f.p = 2.305 nF = 0.13

↑ Integrated flow during liquefaction.

Half amount of LXe in LP corresponds 13% filled in upper level meter.

So we should change recovering tank to second one when level reached that point.

22:00 found that HV had been ON. hvedit ~~does~~ work so we manually turn off.

22:30 Open fully the exit valve

10 Mar 05

23:50 For suppressing MSCB error message
"Error Reading MSCB bus",

MSCB Address [9-16]

in /Equipment /Environment /SETTINGS / Devices /MSCB

in ODB

are changed from "3" → "10"

↑ ↑
address dummy

of -
the Broken module

| Time | Heater (Minco) | Voltage | Flow |
|-------|----------------|-------------|---------|
| 22:00 | | 40V → 35V | 22l/min |
| 22:34 | | 35V → 30V | " |
| 23:08 | | 30V → 25V | " |
| 23:14 | | 25V → 20V | " |
| | | (20V → 25V) | " |

Off the heater for a moment and see the flow and N2 flow

12:30 heater voltage 30V. control on
evacuate outer vessel

11 / March / 2005

15:30. heater voltage was 40V → 30V

16:10. N2 filled to outer vessel. about 1/8 atm.

12 / Mar / 2005

| | PT2 #8 | PT3 | PT4 | PT5 | PT6 | PT7 | PT8 |
|-------|-----------|--------|--------|-----------------------------|--------|-----|--------|
| 11:15 | 70.0 Ω | 68.2 Ω | 71.0 Ω | 78.9 Ω 71.7 Ω | 79.4 Ω | o | 71.6 Ω |
| 12:00 | 70.2 Ω | 67.7 Ω | 70.3 Ω | 71.2 Ω | 77.0 Ω | x | 71.3 Ω |
| 13:10 | 69.8 Ω | 67.7 Ω | 70.7 Ω | 71.0 Ω | 77.9 Ω | x | 71.3 Ω |
| 14:15 | 69.8 Ω | 67.7 Ω | 70.9 Ω | 71.6 Ω | 78.1 Ω | x | 71.2 Ω |
| 15:30 | 70.0 Ω | 67.9 | 71.2 | 71.4 | 78.7 | x | 71.2 |
| 16:45 | 69.8 Ω | 67.7 | 71.0 Ω | 71.7 Ω | 78.7 Ω | x | 70.9 Ω |

13 / Mar / 2005

11:33 add Lg. N2

13:00 Inner pressure starts decrease

| Time | SI top | outlet pump | SI bottom | HLup | inlet pump 1 | Minco heater |
|-------|--------|-------------|-----------|--------|--------------|--------------|
| 14:30 | 12.3 Ω | 70.1 Ω | 73.74 Ω | 74.1 Ω | 82.7 Ω | 83.6 Ω |

14:08 valve close. inner press. 0.122 MPa

14:45 Minco Heater off

Xe tank. pneumatic valve close

15:25 evacuate outer vessel

15:15 HV file 050220_2.hv loaded.

16:19 #9326 pedestal. gas Xe 0.125 MPa.
 16:26 #9327 LED "
 16:36 #9328 alpha. " 50k
 16:45 #9329 " " 100k
 17:09 #9330 " " 100k.
 17:27 #9331 pedestal.
 17:28 #9332 LED
 17:37 #9333 alpha. 160k.

18:10 HT off.
 Xe recovery restarted.

4/Mar/2005

4:30 recovery finished
 - close valves between IV and Liquid pump
 - fill IV with 1.0 atm N₂
 - OV with 0.5 atm N₂
 • Regulator set point 0V
 • Close low pressure valves

9:00 All Computers and Electronics for Power Cut, ~~turn off~~
 power cut pull on tomorrow morning. turn on computers and electronics and CBRP compresses
 12:00 Xe Tanks connect each other.

23:30 Xe Tank pneumatic valves connected pneumatic pipe not via valve controllers but directly pneumatic line for tomorrow power cut.

| | SMtop | Outlet lg. pump | SM bottom | HL up | Inlet lg. pump | N ₂ out | Minco. |
|-------|------------------|----------------------------|----------------------|------------------|---------------------------|-----------------------------|-------------------|
| | SMtop | Outlet lg. pump | SM bottom | HL up | Inlet lg. pump | N₂out | Minco. |
| 16:30 | 73.06 | 70.92 | 74.56 | 74.93 | 84.20 | X | 76.75 |
| 17:00 | 73.24 | 71.12 | 74.70 | 74.70 | 84.29 | X | 76.58 |
| 17:30 | 73.61 | 71.80 | 75.19 | 75.30 | 84.72 | X | 76.63 |
| 18:00 | 74.27 | 72.20 | 75.70 | 76.35 | 85.12 | X | 76.63 |